

Manual PICO-PROFILE with PICO-T3PN Segments







Thank you for purchasing a IMKO instrument.

Should you have any queries please don't hesitate to contact your local distributor or address directly to IMKO:

IMKO Micromodultechnik GmbH Am Reutgraben 2 D-76275 Ettlingen Germany

Phone:	+49-7243-5921-0
Fax:	+49-7243-90856
E-mail:	info@imko.de
Internet:	http://www.imko.de



Manual for TRIME[®]-PICO PROFILE

List of contents

1		Functional Description	5
	1.1		
	1	1.1 Operation mode A (Bus communication)	
	1.2		
	1.3	Installation hints	6
2		Technical Data	7
	2.1	Probe length, spacers and TECANAT access tube	8
	2.2	How to connect PICO-T3PN with connection tube set	9
	2.3	PICO DataLogger cable 5m or 15m length10	C
	2.4	PICO DataLogger cable 25m, 50m or 100m length10	C
	2.5	PICO IMP-Bus cable 5m, 4-pin female connector (Item no.: 300115)	C
	2.6	Accessories - SM-USB (Converter Module), (Item no.: 100020)	1
3		Access Tubes and Augers	2
	3.1	Instructions for Access Tube Installation13	3
	3.2	Inserting and Fixing the Rubber Bung13	3
	3.3	Instructions for Water Proof Installation14	4
4		Remote Power Supply to TRIME PICO-PROFILE	5
5		Datalogging configuration examples10	6
	5.1	PICO probes connected to a PC under Windows16	6
	5.2	PICO probes connected to a SDI12 Datalogger1	7
6		Information on Lightning Protection of a Sensor Network	B
	6.1	Introduction18	8
	6.2	Excess voltage protection on 110/220V mains supply18	8
	6.3	Protection of modem and telephone lines18	8
	6.4	Excess voltage protection for network modules by "SM-Blitz"	8
	6.5	Lightning protection on meteorological towers18	8
	6.6	Installation instructions for SM BLITZ lightning protection modules	8
	6.7	Conclusion	8



7		TRIME PICO-T3PN theory	. 19
	7.1	Introduction	. 19
	7.2	Measuring Field	. 19
	7.3	Measuring experiences	. 20
	7.4	Summary	. 20
8		Basic Calibration with the Calibration Set	. 21
	8.1	What is a basic calibration?	. 21
:	8.2 8.	What are the benefits of the calibration set for the user?2.1Calibration set for TRIME probes	
	8.3		
		.3.1 Preparation of the glass beads.3.2 Basic calibration procedure	
		.3.3 Material specific calibration	
	8.4	Soil Density Correction with PICO Sensors	. 23
9		Quick guide for the Software PICO-CONFIG	. 25
	9.1	Connection of the RS485 or the IMP-Bus to the SM-USB Module	. 25
	9.2	Bus Configuration and Probe Scan	. 26
9	9.3	Scan of connected PICO probes on the serial interface	. 27
9	9.4	Configuration of Measure Mode	. 27
9	9.5	Analogue outputs of the PICO probe	. 28
9	9.6	Simulate Analog Output	. 28
9	9.7	Selection of the individual Calibration Curves	. 29
9	9.8	Test Run in Measurement Mode A	. 30
9	9.9	TRIME WIN MONITOR Datalogger Software	. 31
	9.10	0 Calibration Curves	. 32
	9.11	1 Creating a linear Calibration Curve for a specific Material	. 35
9	9.12	2 Nonlinear calibration curves	. 35
10		Electrical Conductivity with TRIME Probes	. 37
	10.1	1 The analysis of Soils for Electrical Conductivity EC _{TRIME}	. 38
	10.2	2 EC _{TRIME} Measurement Range for TRIME Probes	. 39
	10.3	3 Measurement of Pore-Water-Conductivity or Salt-Content of Soils	. 39
	10.4	4 Creation of a Moisture/ EC _{TRIME} Diagram for a particular Soil	. 42
11		Software Error codes	. 43
	11.1	1 Software Errors which will be coded with 4 digits	. 43
	11.2	2 PICO Errors (Firmware errors)	. 45
12		Savety Notes	. 46



1 Functional Description

The intelligent and compact TRIME PICO-PROFILE sensors are measurement devices for continuous and nondestructive determination of volumetric moisture. They are designed for stationary subterranean field use. A variety of installation options (greater depth, vertical or horizontal orientation) offer a wide range of applications.

TRIME PICO-PROFILE systems require an external 7-24V/DC power supply. They are designed for connection to a data logger or a PC for monitoring and data logging purposes.

Your sensor is supplied ready for use and works in a wide range of standard soils. For further information please check the details under Section 6!

1.1 Operation modes

TRIME PICO-PROFILE systems are supplied with a RS485 or IMP-Bus interface and have no analogue outputs.

TRIME PICO-PROFILE sensors can be easily connected to SDI12 dataloggers (Operation mode A). A detailed description of how to select a specific operation mode for your application can be found below.

1.1.1 Operation mode A (Bus communication)

TRIME PICO-PROFILE systems can be connected directly to the RS485 bus or IMP-Bus (4-wire bus system), either via SDI12 datalogger or via the SM-USB level converter module for use in conjunction with the TRIME WIN MONITOR datalogging software.

If multiple sensors are to be wired as a network, IMKO offers 3-port, 6-port and 12-port distribution modules. Please note that the IMP-Bus cable length and cable diameter must be properly matched as otherwise the energy consumption of the TRIME sensors (100mA @ 12V/DC for 2..3s) may cause a drop in voltage. *More information is available in Section 5.*

Benefits:

- Low power consumption in field installations (with SDI12 datalogger)
- Straightforward installation and simple wiring by virtue of pre-configured standard components (e.g. lightning protection, distribution modules...)
- Systems are supplied ready for use
- Cable length >1000m (with only 4 wires for all sensors)

For use with:

- SDI12 datalogger
- PC under Windows with the software TRIME WIN MONITOR, only with converter module SM-USB.
- IMKO calibration and test software PICO-CONFIG (see page 27, Software download: www.imko.de) only with converter module SM-USB.

1.2 External power supply

In the IMP-Bus the power to sensors can be supplied by battery, a complete solar system or mains power. For serial connection of a large number of sensors (up to 150 sensors) or very long IMP232-bus cables up to 1km it is advisable to use a power amplifier module (SM-23LV) or decentralised power supply.



1.3 Installation hints

Please assure careful installation of the probes with close contact between rod and soil. It is important to avoid air pockets around the rods as the highest measuring sensitivity is directly around them.

Air pockets around the probe rods can reduce the measured moisture reading. Where saturated soils are involved, water-filled air pockets will result in an exaggerated reading.

IMKO supplies pre-drilling sets for an optimal preparation of the installation point avoiding compaction of soil otherwise caused by the insertion of the rods.



2 Technical Data

Technical Data

	Measuring Segment PICO-T3PN
Power supply:	7V24V-DC
Power consumption:	100mA @ 12V/DC during 23sec. of measuring
Moisture measuring range:	0100% volumetric water content
Accuracy (in % volumetric water content):	± 2%
Repeating accuracy/reproducibility:	< 0,1%
Conductivity measuring range:	EC _w 0>20dS/m
Measurement volume:	1dm²
Operating Temperature:	-15°C50°C (extended temperature range on request)
Calibration:	Selectable calibration curves: 1. Standard calibration (soil ±3%) 2. Sand 3. Clay 4. Loam 5. Dielectric Permittivity
Probe body:	waterproof sealed PVC (IP62)
Size:	210 x Ø 40mm
Access tube:	Tecanat Ø44mm, max. 3m length
Connection tube set between two PICO-T3PN segments:	Made of aluminium and PVC. Incl. BCS6 male-connector and female-connector. Available lenghts: 10cm, 30cm and 80cm, special lengths on request!
Interfaces:	IMP-BUS or SDI12 R485 on request
Cable:	2m, 5m and 10m cable with 4-pin female connector



2.1 Probe length, spacers and TECANAT access tube

In the following drawing, the dimensions of the probe PICO-T3PN, the spacers and the TECANAT access tube can be seen. Possible are maximum 4 segments PICO-T3PN at 1 meter tube length, beginning with the first segment direct at the soil surface.





2.2 How to connect PICO-T3PN with connection tube set

Connection tube sets are available in 10cm, 30cm and 80cm.



The measuring segment PICO-PROFILE T3PN is easy to connect with the connection tube set by screwing.





2.3 PICO DataLogger cable 5m or 15m length

Wiring:	
Blank:	Shield
Yellow:	COM (IMP-Bus)
Grey:	R/T (IMP-Bus)
Brown:	GND
White:	+Vs (+12VDC)
Blue:	RS485B or SDI12 Data
Pink:	RS485A

2.4 PICO DataLogger cable 25m, 50m or 100m length

For longer cable length it is necessary to use another cable:

Wiring:		
Pink:	Shield	
Grey:		
Violet:		
White:		=
Blue:	COM (IMP-Bus)	
Green:	R/T (IMP-Bus)	-
Black:	GND	-
Red:	+Vs (+12VDC)	
Yellow:	RS485B or SDI12 Data	
Brown:	RS485A	



2.5 PICO IMP-Bus cable 5m, 4-pin female connector (Item no.: 300115)

Pin layout	:		0 0	
Pin 1:	+Vs (+12VDC)			
Pin 2:	R/T (IMP-Bus)	1	- (2 3) -4	
Pin 3:	GND			
Pin 4:	COM (IMP-Bus)			
List of abb	reviations:			
+Vs: + Volt	age supply (12V/DC)	OM: Common (ground for IMP-Bus)		
GND: Grou	ind (for voltage supply)	T: Receive/Trans	smit (on IMP-Bus)	



2.6 Accessories - SM-USB (Converter Module), (Item no.: 100020)

For connecting of the TRIME PICO-PROFILE (see 3.1 to 3.3, for 3.2 & 3.3 adapter required) via the USB-Interface *to a PC*. The module offers 2 Sensor-Interfaces, IMP-Bus (IMKO specific) and RS485 (industrial standard). One sensor can be powered out of the USB-Interface, if multiple sensors are connected external power supply is required!





3 Access Tubes and Augers

The penetration depth of the measurement field of the 44mm TRIME tube-access-probe is up to 100mm into the soil. The measurement sensitivity is the highest near the access tube and decreases exponentially into the medium. Therefore the insertion method of the access tube is very important.



Pre-boring of bore holes with standard augers destroy the soil texture, because it is difficult to come to a good and close contact of the access tube inside the soil. With the described auger equipment it is possible to set the access tubes directly into homogenous soils without pre-boring. In very stony soils it is not possible to use this method. Therefore it could be possible to use an Edelman-Auger for preboring and closing the air gaps with mud. Changes in soil structure, and a delay time (up to 4 weeks) before it is possible to come to precise measurement values must then be accepted.

The IMKO auger equipment consists of: access tube support pillar with three soil anchors, ramming head with clamp device, screw auger with handle, clamp device for the rubber bung, and an internal steel guide/protection tube. Deliverable is a small version without soil anchors for setting of glass fibre tubes with 1m length.

A steel cutting shoe is glued into the access tube. The screw auger, that moves easily within the guide tube is used to drill out soil to about 0,1m below the cutting shoe. Depending on soil homogeneity, the tube can be hammered 5..10cm into the soil. This cycle is repeated until the tube is fully installed. The internal protection tube is then removed and the access tube can be sealed by a rubber bung.



3.1 Instructions for Access Tube Installation

The following instructions should be taken account of:



The ramming head, the TECANAT access tube and the internal steel guide and protection tube are one unit.

The internal protection tube has to lie on the cutting shoe of the access tube, before it can be fixed with a screw to the ramming head. This fixation is necessary to secure the shoe against being squeezed out by the returning force of the hammer-blow.

The access tube, however, is fixed to the ramming head by the clamp device. Both fixation screw and clamp device should be controlled during the installation process and be re-adjusted with the adjustable levers if necessary.

The screw auger that moves easily within the guide tube is used to drill out soil to about 0,1m below the cutting shoe. Depending on soil homogeneity, the tube is then hammered 5-10 cm into the soil with a plastic hammer (with open support pillar clamps!). The guide tube is not used for making the hole in the first instance because this could result in soil compaction around the hole, which would lead to higher measurement values.

The access tube support pillar with three ground anchors avoids vibrations that would cause air gaps. Then the clamp of the support pillar must be closed and the soil can be removed with the screw auger. When the tube reach the support pillar, it must be removed and the access tube can be inserted to the final

depth without the support pillar.

A plastic collar should be mounted around the tube to prevent water from running down the tube wall and a plastic cap to protect the tube against rain.

3.2 Inserting and Fixing the Rubber Bung

- 1. Orient the rubber bung with the black rubber washer downwards.
- 2. Insert the rubber bung into the TECANAT tube.
- 3. Let it glide down or push to the bottom of the tube.



4. Fix the rubber bung with the screwing adaptor.





The rubber bung can be pushed into the access tube with the screw- adapter and the handle (the screw-adapter can be replaced instead of the screw auger) and can be fixed at the bottom of the access tube with two turns of the handle. If pushing down the rubber bung turns out to be difficult, just apply some talcum powder on the rubber bung sides and into the access tube.

3.3 Instructions for Water Proof Installation

Attention! Risk of Distruction!

The TECANAT tube is made of highly water resistant material in order to warrant for a long lifespan of the tube and the installed PICO-PROFILE probe. In spite of the robust construction, it is to ensure that no water can penetrate into the TECANAT tube. The rubber bang has to be installed correctly and the tube head has to be sealed very secure against water ingress.

Any damage caused by faulty installation is not covered by the warranty!



2017-05-30

4 Remote Power Supply to TRIME PICO-PROFILE

The operation of TRIME sensors may cause problems when power has to be supplied via long cables. There are limitations to the maximum cable length depending on the cable diameter.

When power is supplied over long distance the maximum cable length depends on the cable cross section A, the supply voltage V_s and the number n of the sensors measuring simultaneously. Device-specific data also be applied to the formula:

Power consumption during measurements:	I _{norm}	=	100mA @ 12V/DC
Power consumption at min. voltage:	I_{max}	=	175mA @ 7V/DC
Supply voltage:	V_{s}	=	12V
Minimum sensor voltage at circuit end:	V_{min}		
Wire cross section:	А	=	0,34mm²
Specific electrical resistance of copper:	ρ	=	$0.0178\Omega x mm^2 / m$
Number of sensors:	n	=	1

The maximum possible circuit length I_{max} can then be calculated in the following manner:

$$l_{\max} = \frac{A \cdot (V_s - V_{\min})}{2 \cdot \rho \cdot n \cdot I_{\max}}$$

Please see the following the following example:

In the IMP232 environmental measurement system a bus cable with a wire cross section of A = 0.34 mm^2 is normally used. We further assume that the power supply voltage is V_s =12 V and only one sensor is designated to measure. Thus n = 1.

$$l_{\max} = \frac{0.34mm^2 \cdot (12V - 7V)}{0.0356\Omega \frac{mm^2}{m} \cdot 1 \cdot 0.175A} = 270m$$

In the above calculation, no tolerance is included; for security reasons the calculated cable length should be reduced by 10% to obtain a realistic value.

In order to increase the maximum possible cable length several solutions are feasible.

- Using cables with larger conductor diameters By using 6-core conductor cables instead of 4-core, the cable length can be doubled as two extra cores can be used for power supply. Cables with conductors of larger diameters will further increase the maximum cable length possible.
- Increasing the power supply voltage
 Power supply voltage can be increased up to 17V, thereby raising the maximum length from 270m to 540m in the example calculation above.
- Installation of buffer batteries in the distributor Additional storage batteries close to the TRIME sensors, e.g. in the distributor, allow cable lengths up to 1km and enable simultaneous measurement of several sensors. However, this method requires an additional charging circuit for the buffer storage battery.
- Installation of a voltage regulator at the distributor
 Voltage loss in the cable can be reduced with a 30V power supply and an installation of a voltage regulator directly in front of the TRIME sensor, thus allowing circuit lengths of up to 1km.

Which solution is best suited mainly depends on the nature of the power supply of the measurement system:

- Battery supply: solution 1 and possibly solution 3 should be considered, the latter being relatively expensive.
- Mains supply: solutions 1 and 2 could be combined, or, more expensive, solutions 2 or 4 could be chosen.



5 Datalogging configuration examples

PICO probes offer the possibility to use three different options for serial communication:

- The robust and secure (opto-isolated) IMP-Bus for large area monitoring networks with cable length >500 meter,
- the RS485 as an international standard serial interface for connection of PICO probes to standard PC's, PDA's and other computers,
- or the SDI12 interface for direct connection to SDI12 dataloggers.

5.1 PICO probes connected to a PC under Windows

With IMKO's small software package TRIME WIN-MONITOR (cost free) it is possible to collect measurement data from up to 60 PICO probes to a PC. TRIME WIN-MONITOR runs under Windows. With the software PICO-CONFIG it is possible to configurate single PICO probes, e.g. to select a special calibration curve inside the PICO probe. The connection to the PC can be configured via the module SM-USB and a USB-port to the PC.



Standard PC/Laptop under Windows. For datalogging with TRIME WIN MONITOR (cost free) and PICO CONFIG (Probe Configuration).





5.2 PICO probes connected to a SDI12 Datalogger



6 Information on Lightning Protection of a Sensor Network

6.1 Introduction

Lightning strikes can cause considerable and costly damage to unprotected electronics. The equipment is often totally destroyed. A good number of users are not or only partially insured. Customers who have lightning protection insurance must comply with defined clauses regarding lightning and excess voltage. Insurance companies only cover the damage when compliance with the defined clauses has been proven. IMKO strongly recommends adequate lightning / excess voltage protection equipment for environmental measurement systems.

6.2 Excess voltage protection on 110/220V mains supply

Lightning strikes in proximity to high-voltage transmission lines can cause excess voltage in the mains power supply which may result in damage of electronic components. Environmental measurement systems with 110/220V mains supply are at risk from this excess voltage. It may affect the whole system through the power supply unit and the central station (SDI12 datalogger or SM-USB). Excess voltage can even enter the measuring system through the data acquisition computer's mains power supply. An excess voltage protection is highly recommended for all 110/220V devices connected to a sensor system

6.3 Protection of modem and telephone lines

Telephone lines are at risk from excess voltage. If a modem is connected to the measurement system the telephone line should also be protected by a lightning protection module.

6.4 Excess voltage protection for network modules by "SM-Blitz"

Excess voltage caused by lightning strokes in close proximity to the environmental measurement test system may enter the IMP-Bus transmission lines. Longer lines increase the risk of lightning strikes. Theoretically maximum protection is achieved by installation of a lightning protection module (SM-Blitz) in front of each SM-Module. Lightning protection is not cheap but it is certainly worthwhile. A compromise should be found between costs and the maximum-affordable protection, i.e. interconnection of adjacent SM-Modules to lightning protected groups.

6.5 Lightning protection on meteorological towers

SM-Modules installed on meteorological towers cannot be protected from lightning strikes. The field strength resulting from the electromagnetic fields and the associated accumulated energy will cause damage to the electronics. Two solutions to the problem:

- Erect a higher lightning conductor close to the meteorological tower serving as a lightning conductor.
- Install the measuring modules a number of metres away. Then all lines coming from the tower have to be protected by lightning protection modules.

6.6 Installation instructions for SM BLITZ lightning protection modules

Basically, there are two potential sources of risk in the field of environmental measurement technology: transmission lines and sensors or network devices. Lightning protection modules should always be installed at the beginning and at the end of a circuit in order to protect the electronics from excess voltage (Attention: SM-Blitz modules have a protected and an unprotected side).

The SM-BLITZ lightning protection module has to be grounded using a ground conductor with a wire crosssection of at least 6 mm² screwed to the long side of the module. A 2-metre long grounding rod may serve as a ground conductor. Grounding is optimal when the grounding rod is in direct contact with ground water.

6.7 Conclusion

Only limited protection against excess voltage is possible where natural phenomenon such as lightning strikes are concerned. Direct lightning strikes may cause damage nevertheless.



7 TRIME PICO-T3PN theory

7.1 Introduction

The measuring of soil water content with Time Domain Reflectometry is now a well established method. However water content profiling is not possible with conventional TDR rod probes. The TRIME tube probe was developed for this reason.

Since 1994 the TRIME-T3 has found numerous applications in earth and environmental sciences, fulfilling even the most exacting requirements.

7.2 Measuring Field

The effective penetration depth of the probe is about 10 cm with the highest sensitivity in the immediate vicinity of the access tube and decreases exponentially with distance. Figure 1 shows the electric field distribution of the probe and the approximate measuring volume.



Figure 1: Electric field distribution of the TRIME probe and approximate measuring volume.

Note that the necessity of a close contact between access tube and material is vital for reliable measurements and that the tubes should be installed by our recommended method.

For example at an assumed water content of 15 vol. % an air gap of 1 mm around the whole length of the tube would result in an underestimation of 1 - 2 vol. %.

At a water content of 25 vol. % the error would be 5 vol.-% with an air gap around the tube.

At very high water contents (50 vol. %) errors may reach 10 vol. % with an air gap around the tube.

In the case of a water filled gap under conditions of saturation the gap error would be much smaller.

Problems may arise, however, in very inhomogeneous soils and when drilling under very dry conditions. For these soils other drilling methods are recommended (e. g. pre-boring with an Edelman auger, washing mud into the cavity around the tube). Losses in accuracy must then be accepted, and measurements immediately after installation are not recommended.

Problems can also arise in swelling and shrinking soils, since cracks develop especially along the access tubes.



7.3 Measuring experiences

The new TRIME technique was thoroughly tested in the field and compared both to neutron probe measurements and thermo gravimetrically determined values.



Figure 2: Comparison of TRIME measurements and gravimetric water content determination for a clayey soil.

7.4 Summary

The TRIME tube probe is a promising new tool for determining water content profiles with the TDR method. Fast, routine and nondestructive measurements of water content without the use of hazard-ous radioactive materials are possible. A measuring accuracy of ± 2 vol.-% is possible, provided that soil and access tube are in close contact and bulk soil electrical conductivity doesn't exceed 2 dS/m.



8 Basic Calibration with the Calibration Set

8.1 What is a basic calibration?

Basic calibration serves to compensate the cable length and tolerances of the probe mechanics (thickness of the rod coating, rod length, etc.). After two measurements, one in dry and one in water-saturated glass beads, the calibration data is calculated and stored in the TRIME sensor.

Every TRIME PICO-PROFILE sensor must be calibrated before it can supply proper measurement results. **Basic calibration is carried out by IMKO in the factory prior to shipment.**

8.2 What are the benefits of the calibration set for the user?

With the calibration set you can easily carry out basic calibration of your TRIME sensor yourself.

If defective probe rods must be changed, you can perform the required basic calibration yourself.

The calibration set <u>cannot</u> be used for establishing a material (soil) specific calibration. For this purpose a measurement dataset must be created for the specific material. The complementary calibration program *TRIME-Tool* is required to calculate the calibration data for this dataset and to download it to the TRIME PICO-PROFILE-Probe.

8.2.1 Calibration set for TRIME probes

For basic calibration of TRIME probes.

- 2 x boxes (7 litres.)

- 22kg glass beads





8.3 How to perform a basic calibration?

8.3.1 Preparation of the glass beads

The glass beads, supplied with the calibration set, have to be prepared first:

Fill up one bucket until the rods of the probe can be immersed completely. To obtain a consistent density, knock the bucket on the ground several times.

The density of the glass beads increases with frequent insertion of probes. Therefore the glass beads should be poured out into another bucket and poured back to achieve the original density.

Now the second bucket has to be filled with water in order to be able to fill in the glass beads without leaving air-bubbles. An additional precaution against air-bubbles is to stir slightly while filling in the glass beads. The container must now be knocked on the ground several times to obtain a consistent density. The surplus water must be poured out until the depth of the water film above the glass beads is less than 2mm.

The water-saturated glass beads should be in a temperature range between 20°C and 25°.

- <u>Attention</u>: Water dissolves Na₂O and K₂O from glass which causes a rising pH-value and higher electrical conductivity. New glass beads have to be washed intensively with tap water!!!
 - 1. Fill a bucket with water
 - 2. Stir the beads under water to drive out all air bubbles
 - 3. Pour out the water. This procedure should be done with new glass beads at least five times, each time with fresh water. If the glass beads have been in use for a prolonged period, three times is sufficient.

Please note that the electrical conductivity of the water-saturated glass beads medium increases already after a few days storage. Therefore the glass beads must be washed again before the next calibration.

8.3.2 Basic calibration procedure

Basic calibration must be performed using the calibration program *Pico-Config*. Please read the information about basic calibration with *Pico-Config* in **chapter 12** of this document.



Fig.: PICO T3 probe head inserted in a TECANAT calibration tube which again is installed in dry glass beads. Make sure that the tube probe's wave guides are completely inserted into the glass beads when measuring. The basic alignment will fail if the wave guides are in part surrounded by air.

8.3.3 Material specific calibration

Your TRIME measuring system operates with a universal calibration for mineral soils as a standard.

The following parameters limit the application range of the universal calibration:

- Clay content: >50%
- Organic content: >10%
- Bulk density: <1.1kg/dm³ or >1.7kg/dm³

Exceeding these limits may cause the tolerances given on page 5 to be overstepped.

Material-specific calibration is advisable if your soil is listed above or if you require accuracy down to the last digit. The **PICO-CONFIG** software is required for setting up a material-specific calibration (*download under* **www.imko.de**).

A test series with reference values is necessary for performing material-specific calibration (e.g. Oven drying at 105°C until weight is constant). The test series –and consequently the calibration– should include minimum and maximum moisture values. TRIME readings and reference values are compared in a table. The calibration coefficients must then be calculated and uploaded to the TRIME-device.

8.4 Soil Density Correction with PICO Sensors

One of the most important advantages of PICO sensors is, that it is possible to measure volumetric moisture for all soil types with <50% clay and <10% organic with just one calibration. The only factor is the offset caused by the differing soil density, which can be compensated by a formula which IMKO offers at its homepage in an Excel-File **"Soil Density Correction"**.

Volcanic soils have normally lower densities, clayey soils can have higher densities.

It is possible to make an offset correction with Excel functions after datalogging of the stored moisture values. But it is also possible to change the offset calibration parameter m0 inside the PICO probe before installation and use of the PICO sensor. With help of the software tool PICO-CONFIG it is possible to load the "**Universal Soil**" calibration curve of the PICO sensor, to change it and store it back to the sensor.

Necessary is to change the parameter m0 of the "Universal Soil" calibration curve inside the PICO probe. If a soil has a lower density as 1,4 than the PICO senor measures a too low moisture value and therefore the calibration parameter m0 has to be increased with an offset correction value.

E.g. if the soil density is 1,3 than the calibration parameter m0 has to be increased with +1,21. If e.g. the calibration parameter m0 = -30 inside the PICO sensor than it has to be increased to m0 = -30 + 1,21 = -28,79

If the soil has a density higher than 1,4 than the calibration parameter m0 of the "**Universal Soil**" calibration curve has to be decreased with the appropriate offset correction value.



	Soil Density	Factor		Error in % due to deviating Density	m0 Offset Correction Value
-0,3	1,10675	12,12	-17,05	-3,64	+3,64
-0,2	1,20675	12,12	-17,05	-2,42	+2,42
-0,1	1,30675	12,12	-17,05	-1,21	+1,21
0	1,40675	12,12	-17,05	0,00	DEFAULT
0,1	1,50675	12,12	-17,05	1,21	-1,21
0,2	1,60675	12,12	-17,05	2,42	-2,42
0,3	1,70675	12,12	-17,05	3,64	-3,64
0,5	1,90675	12,12	-17,05	6,06	-6,06
0,6	2,00675	12,12	-17,05	7,27	-7,27
0,7	2,10675	12,12	-17,05	8,48	-8,48

The correction values for some soil density values are outlined below:

Load the Excel-file **"Soil Density Correction"** from IMKO's homepage **"Support"** and **"PICO-CONFIG"** and enter your soil density for getting the appropriate offset correction value.

Cal	Act	CallD-P	CalName in Probe	MatID-P	TemID-P	DenID-P	Set Active Calib
0		00000	No Calibration	00000	00000	00000	Default Calibration Item
1	Α	02002	Universal Soil,Vol	02001	02000	02000	1
2		02017	Low density soil (1.1kg/dm³)	02017	02000	02000	, , , , , , , , , , , , , , , , , , , ,
3		02018	High density soil (1.7kg/dm³)	02018	02000	02000	Set Default Calib
4		02003	Dielectric coefficient	02002	02000	02000	
5		02024	Sand D=1.4 grav.	02024	02000	02000	Calibration Name
6		02019	Sand D=1.5 grav.	02019	02000	02000	Universal Soil,Vol
7		02025	Sand D=1.6 grav.	02025	02000	02000	Of livers al Soli, vol
8		02020	Lightly-Sand moisture (grav.)	02020	02000	02000	Set
9		02029	Sand D=1.5 in Bucket (grav.)	02029	02000	02000	Jet
10		02027	Gravel 2-8mm (grav.)	02027	02000	02000	
11		02028	Gravel 8-16mm (grav.)	02028	02000	02000	Material Coeffs — Temp Coeffs —
12		02030	Quarz Sand (grav.)	02030	02000	02000	m0 -30.1336 t0 20
13		02031	Feldspar (grav.)	02031	02000	02000	
14		02021	Peat	02021	02000	02000	m1 0.525621 💦 t1 0
15		02026	1/10tp	02026	02000	02000	m2 -0.0032093
							m3 9.81034e-0 3 0
							m4 -1.36253e-(4 0
							m5 7.164e-012 5 100
							Set Set
							Save
							Read Read

9 Quick guide for the Software PICO-CONFIG

9.1 Connection of the RS485 or the IMP-Bus to the SM-USB Module

The SM-USB provides the ability to connect a PICO probe either to the standard RS485 interface or to the IMP-Bus from IMKO, which enables the download of a new firmware to the PICO probe. Both connector ports are shown in the drawing below.

The SM-USB is signalling the status of power supply and the transmission signals with 4 LED's. When using a dual-USB connector on the PC, it is possible to use the power supply for the PICO probe directly from the USB port of the PC without the use of the external AC adapter.



Please note: The PICO-PROFILE wiring and cabling has other colours!

How to start with the USB-Module SM-USB from IMKO

- Install USB-Driver from USB-Stick.
- Connect the SM-USB to the USB-Port of the PC and the installation will be accomplished automatically.
- Install Software PICOConfig-SetUp.msi from USB-Stick.
- Connection of the PICO probe to the EX9531 via RS485A, RS485B and OV.
- Check the setting of the COM-Ports in the Device-Manager und setup the specific COM-Port with the Baudrate of 9600 Baud in PICO-CONFIG with the button "Bus" and "Configuration" (COM1-COM15 is possible).
- Start "Scan probes" in PICOConfig.
- The PICO probe logs in the window "Probe List" after max. 30 seconds with its serial number.

Note 1:

In the Device-Manager passes it as follows: Control Panel \rightarrow System \rightarrow Hardware \rightarrow Device-Manager





COMx set must be between COM1....COM9 and it should be ensured that there is no double occupancy of the interfaces.

If it comes to conflicts among the serial port or the USB-SM has been found in a higher COM-port, the COM port number can be adjusted manually:

By double clicking on "USB Serial Port" you can go into the properties menu, where you see "connection settings" – with "Advanced" button, the COM port number can be switched to a free number.

Eigenschaften von USB Serial Port (COM4)	2 🔀	Enweiterte Einstellungen	tür COMI				2
Mgemein Anschlussindslungen Tiebe	r Detaks	-	COM-Anschlusstummer:	COMH	2		OK	_
Bits pro Sekunde Deterbits			Utill Packetprofien Reduzieren Sie die Werte, u Entohen Sie die Werte für e	COML (berets belegt) COM3 (berets belegt) COM4	Baudra	ten zu beheben.	Abbred	hen
Partia: Stoppbly:	Contraction -		Empfangen (Sytes): Senden (Sytes):	COMP (benetis belegt) COMP (benetis belegt) COMP (benetis belegt) COM10 (benetis belegt) COM11 (benetis belegt)				
Russteuerung	Keine		OMEnstellungen Reduzieren Sie die Werte, u	COML2 (benets belogt) COML3 (benets belogt) COML4 (benets belogt) COML5 (benets belogt)	Alterty	ne Optionen / für serielle Schnittstelle		P
En	witen Wiedshen	telen	veringern. Wartspet (ne):	COM16 (bereits belegt) COM17 (bereits belegt) COM19 (bereits belegt) COM19 COM29 (bereits belegt)	Serieller Abbred	Drucker ten der Kommunikation, withaltet wird	enn das Geråt	
			Taxenda	C0421 C0422	Event b Geräck	ei unvorhergesehener Ent		
-			Minimale Ancahle der Lose- (ms): Minimale Ancahle der Schre (ms):	COM25 (benaits belegt)	Abechal	Hießen der Verbindung R1 ben der Modernansteuerut vren des Gerähs		0
	OK Ab	beachen		COM58				

After changing the COMx port settings, PICO-CONFIG must be restarted.

9.2 Bus Configuration and Probe Scan

With PICO-CONFIG it is possible to make process-related adjustments of individual parameters of the PICO probe. Furthermore the measurement values of the PICO probe can be read from the probe via the RS485 or IMP-Bus interface and displayed on the screen.

In the menu **"Bus**" and the window **"Configuration**" the PC can be configured to an available COMx-port with the Baudrate of 9600 Baud.



	Bus Configuration	×
PicoConfig	General	
Exit Bus Help Configuration Scan Probes	Serial Port COM1 COM1 COM2 COM3 COM4 COM5 COM6 COM6 COM7 COM9 COM9 COM10	el

9.3 Scan of connected PICO probes on the serial interface

In the menu **"Bus**" and the window "**Scan Probes**" the RS485 bus can be scanned for attached PICO probes (takes max. 30 seconds).

PICO-CONFIG reports founded PICO probes with its serial number in the window "Probe List".

No.	SerialNo	ResetBaudrate	ProbeName	HWVersion	FWVersion
	30736	9600	TRIME-PICO	1,12	1.120312
	43	9 -02000-00		2008-075279	
_					
-					
_					
1					

9.4 Configuration of Measure Mode

In "**Probe List**" with "**Config**" and "**Measure Mode & Parameters**" the PICO probe can be adjusted to the desired mode A, B or C (see Chapter "Configuration Measure Mode").



Measure Mode &	Parameters	×
Меа	asure Mode	
A	OnRequest 🥂	E C
Sleep Time in	n Mode C(s) 10	
Wait Time in	Mode B(s)	
	Set	
	Close	
6	LIUSE	

9.5 Analogue outputs of the PICO probe

In the menu "**Config**" and the window "**Analog Output**" the analogue outputs of the PICO probe can be configured (see Chapter "Analogue outputs..").

	Output Channels Moist,Temp C Moist,Condu	ict
Moisture Range Maximum 100 Set Minimum 0. Set		Conductivity(mS/cm)
Minimum JUSet	Minimum -40 Set	

The conductivity range can be selected, depending on the specific soil which is measured. 0..20 dS/m should be selected for soils with a very high salinity content.

9.6 Simulate Analog Output

The PICO PROFILE has no analog outputs! The following text refers to the probes PICO64/32.

By pressing the button **"Simulate"** in the window **"Analog Output**" the analog output 1 and 2 can be simulated in 5 ranges from 0 to 100%. The %-value can be entered directly by hand and the analog output follows by pressing the button **"Set**".



nulate Analog (Dutput	2
	0100.0	0-1V:
Moisture	0	00V 250,25V
Temperature	0	500.5V 750.75V
		1001V
	Set	
		N
	Close	43

9.7 Selection of the individual Calibration Curves

In the menu **"Calibration**" and the window **"Material Property Calibration**" the calibration curves CAL1 to Cal15 which are stored in the PICO probe are loaded and displayed on the screen (takes max. 1 minute). With the mouse pointer individual calibration curves can be activated and tested with the PICO probe by activating the button **"Set Active Calib**". Furthermore, the individual calibration curves CAL1 to Cal15 can be adapted or modified with the calibration coefficients (see Chapter "Creating a linear calibration curve").

A 00000 00000 No Calibration 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000	Cal	Act	-7	CalD-P	ColDF	CalNome in Frobe	CalName in File	ManDP	MatD-F	TemiDP	TenIDF	DenID-P	DenIDF		
C0004 C0000 UmSalding D=1.0. C0001 C0000 C2001 C00000 C2001			-	000000	00000	No Calibration		00000	00000	00000	00000	00000	00000		
C0005 C00000 Un-Sold.og., D-1.2. C0201 C0000 C2002 C0000 C2002 C0000 C2002 C0000 C2003 C00000 <thc2003< th=""> <thc2003< th=""> <thc2003< th="" th<=""><th></th><th>A</th><th></th><th>\$102002</th><th>00000</th><th>Universal SoltVol</th><th></th><th>102001</th><th>00000</th><th>02000</th><th>00000</th><th>02000</th><th>00000</th><th></th><th></th></thc2003<></thc2003<></thc2003<>		A		\$102002	00000	Universal SoltVol		102001	00000	02000	00000	02000	00000		
C0005 C00000 Un-Sold.og., D-1.2. C0201 C00000 C2002 C0000 C2002 C0000 C2002 C0000 C2002 C0000 C2002 C0000 C2003 C0000 <thc2003< th=""> <thc2003< th=""> <thc2003< th="" th<=""><th></th><th></th><th></th><th>02004</th><th>00800</th><th>UniSold g. D=1.0</th><th></th><th>62001</th><th>00000</th><th>02000</th><th>00000</th><th>02001</th><th>00000</th><th></th><th></th></thc2003<></thc2003<></thc2003<>				02004	00800	UniSold g. D=1.0		62001	00000	02000	00000	02001	00000		
02007 00000 Un+Sist dr.g. D=1.6. 02001 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 000000 02008<				02005	000000	Uni-Solld: gr. D=1.2		62001	00000	02000	00000	02002	00000		
02008 00000 U# Said ag. p-1.8. 02001 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02006 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 00000 00000 00000 00000 000000 000000 000000<				02006	00000	Uni-Soil.dt.grD=1.4		02001	00000	02000	00000	02003	00000		
Q2009 Q0000 Um-Solid gr, D=2.0. E2001 Q0000 Q2006 Q0000 Q2006 Q0000 Q2006 Q0000 Q2007 Q00000 Q2007 </th <th></th> <th></th> <th></th> <th>02007</th> <th></th> <th></th> <th></th> <th>02001</th> <th>00000</th> <th>02008</th> <th>00000</th> <th>02004</th> <th>00000</th> <th></th> <th></th>				02007				02001	00000	02008	00000	02004	00000		
02010 00000 U#Sold g.D=22 02001 00000 02007 00000 02007 00000 02007 00000 02007 00000 02007 00000 02008 00000 02008 00000 02008 00000 02008 00000 02009 00000 02009 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 00000 02008 000000 02008				02008				02001	00000	02000	00000	02005	00000		
OC0111 OC000 U+Fold if y, D=24 OC001 OD000 OC000 OC000 <th></th> <th></th> <th></th> <th>02009</th> <th></th> <th></th> <th></th> <th>02001</th> <th>00000</th> <th>02000</th> <th>00000</th> <th>02006</th> <th>00000</th> <th></th> <th></th>				02009				02001	00000	02000	00000	02006	00000		
0 02012 00000 U+Sald g. D=2.6. 02001 00000 02000 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 02009 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000 000000				02010	00800	Uni-Sail.dt.grD=2.2		02001	00000	02000	00000	02007	00000		
1 02013 00000 U+Seit.dr.gr. D=2.8. 02001 00000 02000 00000 02010 00000 02010 00000 02010 00000 02011 00000 02011 00000 02001 00000 02001 00000 02001 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000 000000 <th></th> <th></th> <th></th> <th>02011</th> <th></th> <th></th> <th></th> <th>02001</th> <th>068800</th> <th>02000</th> <th>00600</th> <th>02008</th> <th></th> <th>16367</th> <th></th>				02011				02001	068800	02000	00600	02008		16367	
2 02014 00000 Uni-Solding. D=30. 02001 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000 000000	0			02812	00000	Uni-Saild: gr. D=2.6		02001	00000	02000	00000	02009	00000	Calibratio	n flem
3 02001 00000 Pseudo-time tape 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 02000 0200000 0200000	1			02813	00000	Uni-Soil dt gr. D.=28.		02001	00000	02000	00000	02010	00000		
3 00001 00000 Pseudo-time tp(pe) 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000 000000	2			02014	00000	Uni-Soldigr.D=3.0		10050	00000	02000	00000	02011	00000	Set	Active Calb
5 02000 00000 No Calibration 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 02000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 00000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000000 000				02001	00800	Pseudo-time tp(ps)		02000	00000	02000	00000	02000	00000		
Setficients Mat in Probe Mat in File Tem in Probe Tem in File Den in Probe Den in File Coefficient Correction 0.0.0.0 0.5001 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 CalName CalName Universal Sol/ m0-m5 30.1338/sec001 0.00000e+000 0.000000e+000 0.000000e+000 <				02001	000000			62000		02000	00000	02000	00000	Default	Calibration Item
Deficients Mat in Probe Mat in File Tem in File Den in File Den in Probe Den in File Coefficient Correction 00.00.00 0.3003335ee001 0.000000ee000 2.000000ee000 0.000000ee000 0.000000ee000 Coefficient Correction 11.01 5.256210e.001 0.000000ee000 m0-m5 30.1336 30.1336 3.0100000ee000 0.000000ee000	5			02000	000000	No Calibration		02000	00000	02000	00000	02000	00000		Statesono VIV
Carticianti Mat in Probe Mat in File Term in File Den in File Den in Probe Den in File 00.00.01 -3013355ee001 -0.00000ee000 2.00000ee000 -0.00000ee000 -0.000000ee000 -0.00000ee000 -0.00															
1/1.d1 5.256210e.001 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 m0-m5 30.1336 2/2.d2 3.20332e+003 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 m0-m5 30.1336 3/3.d3 3.810338e+006 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 10-m5 30.1336 4/4.d4 -1.562530e+08 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 10-m5 20 5/5.d5 7.164000e+012 0.00000e+000 1.00000e+002 0.000000e+000 0.000000e+000 0.000000e+000 10-m5 20	oef	cients	1	N	lat in Probe	Mat in File	Tem in Probe	TerrinP	ie .	Den in Prob	e	Den in File		COSINCE	in conceasi
22,22 -3,29332e-003 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 m0-m5 -30.1336 313,43 9.810338e-006 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 10-m5 -30.1336 313,43 9.810338e-006 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 10-m5 -30.1336 35,45 7.16400be12 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 0.00000e+000 10-m5 20														CaName	Universal Soil?
A3/3.43 9.810336e-006 0.000000e+000 0.0000000000														-	00.4000
H H (d 4 -1.362530e-008 0.000000e+000 0.00000e+000 0.000000e+000 0.0000000000														m0-m5	30.1336
15/5/25 7.164000e-012 0.000000e+000 1.000000e+000 0.000000e+000 0.000000e+000 0.000000e+000														14.5	20
														10112	1-2
Set	15,5	d5		7.11	4000e-012	0.0000000+000	1.000000e+002 0	000000e+0	00. 01	LU0000e+00	0.00	0000+000			14100 1
									100				-		Set
				-							-				

The desired and possibly altered calibration curve (Cal1. .15) which is activated after switching on the probes power supply can be adjusted with the button "**Set Default Calib**".

The coefficients m1 = 0.0581 and m0 = -4.05 for individual calibration curves can be entered and adjusted directly by hand and are stored in the probe by pressing the "**Set**" button.



9.8 Test Run in Measurement Mode A

In the menu "**Test**" and the window **Test in Mode A**, the measured moisture values "**Moist**" of the PICO probe are displayed on the screen and can be parallel saved in a file.

Attention: for a test run in mode A it must be ensured that the PICO probe was also set to this mode. If this is not assured, the probe returns zero values.

		N. 11			R 10	13		0.00	1	<u> </u>		<u>s av</u>	2 J.	_
	Conduct	MatTemp	TDRLevel	Moint	MatMat	StdMat	ip:		ASICTemp	ASIECount	Count	Date	Time	10.
	0.000	23.0	98,300	0.000	42.010	-20.000	29.608	1423.848	51.10	847.546	847.546	21.10.2010	16.32.23	
	1.920	23.1	76.200	0.000	-42.010	-20.000	-87.057	1183.255	50.51	704.333	704 333	21.10.2010	16:32:18	
	0.920	23.0	83.900	13,660	13,663	340.3		2942.603	49.98	1751.584	1751.584	21.10.2010	16:32:14	
	1.050	23.0	82,900	10.980	10.980	305.0		2797.243	49.63	1665.059	1665.059	21.10.2010	16.32.09	
	1.050	23.0	82,900	0.000	16.633	31.031	31.031	1669.341	49.10	993.675	993.675	21.10.2010	16.32.05	
	2.220	23.0	73.900	0.000	-42.010	-20.000	-34,218	1400.757	48.80	833.901	833.801	21.10.2010	18.32.01	
	0.530	23.0	85,900	22,020	22.028	430.6		3314.294 3341.618	48.56	1972.834	1972.834 1989.098	21 10 2010 21 10 2018	16.31.56 16.31.52	
Measure	0.460	23.0	87.400	23.060	23.054	441.1		3357.376	46.85	1998.478	1998.478	21.10.2010		
	0.000	23.2	97,100	6.950	6.951	232.7		2499.797	45.04	1488.004	1488.004	21.10.2010	16:31:43	0
and in								a restrict.	10.01				19191111	-
Sime														
Close														
						-								
						-								
						-								

Following measurement values are displayed on the screen:

- Moist Moisture Value
- MatTemp Temperature
- Conduct Bulk Electrical Conductivity EC_{TRIME}
- TDR-Level (for special applications)
- tp Radar time which corresponds to the respective moisture value.

The other values like Count, ASICCount, etc. are for special use.

By clicking **"Save"** the recorded data is saved in a text file in the following path: **PICO-CONFIG.exe-Pfad\MD\Dateiname**

The name of the text file **TestMeas+SN+yyyymmddHHMMSS.sts** is assigned automatically with the serial number of the probe (SN) and date and time.

The data in the text file can be evaluated with Windows-EXCEL.

Basic Balancing in Glass Beads

PICO probes are identical and manufactured precisely. After an exchange of a probe rod it is nevertheless advisable to verify the calibration and to check the basic calibration and if necessary to correct it with a "Basic Balancing".

With a "Basic Balancing" two reference calibration measurements are to be carried out with known set-points ("RefValues"). For the reference media, dry and wet glass beads should be used. **Attention:** Before performing a "Basic Balancing" it must be ensured that the PICO probe was set to "Measure Mode" A. If this is not assured, the probe returns zero values. After a "Basic Balancing" the PICO probe has to be set to "Measure Mode" B or C again.



In the menu **"Calibration**" and the window "**Basic Balancing**" the two set-point values of the radar time **tp** are displayed with 145ps and 625ps.

Reference set-point A: **tp**=145ps in dry glass beads.

The first set-point can be activated with the mouse pointer by clicking to No.1. By activating the button "**Do Measurement**" the PICO probe determines the first reference set-point in dry glass beads. In the column "MeasValues" the measured raw value of the radar time **t** is displayed.

Reference set-point B: **tp**=625ps in wet glass beads. The PICO probe head has to be completely covered with wet glass beads. The second set-point can be activated with the mouse pointer by clicking to No.2. By activating the button "**Do Measurement**" the PICO probe determines the second reference set-point in wet glass beads. In the column "MeasValues" the measured raw value of the radar time **t** is also displayed.

By activating the button "*Calculate Coeffs*" and "*Coeffs* \rightarrow *Probe*" the alignment data is calculated automatically and is stored in the PICO probe non-volatile. With a "Test run" (in Mode A) the radar time **tp** of the PICO probe should be now 145ps in dry and 625ps in wet glass beads.

No	RefValu	ies(tp)/ps	MoistValues(%)	MeasValues(t)/ps	Comments
1		145.0	2.8		dry glass bead
2		625.0	43.9		wet glass bead
Cool	ficients		<calculated></calculated>	<in file=""></in>	<in probe=""></in>
000	ьо		Calculated>	XIIII IIEZ	-374.513
	Ь1				0.242937
			dry.wet	glass bead Do Measurement Calculate Coeffs Coeffs> Probe	

9.9 TRIME WIN MONITOR Datalogger Software

TRIME WIN MONITOR is a datalogger software for measurement data collection of TRIME sensors via the module SM-USB. TRIME WIN MONITOR runs under Windows.

The software and the manual can be downloaded from IMKO's homepage under the menu item "Support", "Software" and "TRIME WIN MONITOR".



9.10 Calibration Curves

PICO probes are supplied with different calibration curves.

Call Universal Soil Vol. is a nonlinear calibration curve 5th order which is suitable for most soils.

Cal 2 to Cal7 (see next pages) can be selected for special applications. These are the calibration functions installed to the upgrade version of the TRIME PICO-PROFILE32/64 delivered together with the measurement of the Radar based Conductivity (RbC). If you want to have these Calibration functions for your current TRIME PICO-PROFILE please contact IMKO GmbH (info@imko.de)

A maximum of 15 different calibration curves (Cal1 ... Cal15) can be stored inside the PICO probe and can optionally be activated via the program **PICO-CONFIG**.

A preliminary test of an appropriate calibration curve (Cal1. .15) can be activated in the menu **"Calibration"** and in the window **"Material Property Calibration"** by selecting the desired calibration curve (Cal1...Cal15) and with using the button **"Set Active Calib**". The finally desired and possibly altered calibration curve (Cal1. .15) which is activated after switching on the probes power supply will be adjusted with the button **"Set Default Calib**".

Nonlinear calibrations are possible with polynomials up to 5th grade (coefficients m0 to m5).

IMKO publish on its website more suitable calibration coefficients for different materials. These calibration coefficients can be entered and stored in the PICO probe by hand (Cal14 and Cal15) with the help of PICO-CONFIG.

Next page:

The following charts (Cal.1..15) show different selectable calibration curves which are stored inside the PICO probe.

On the y-axis the gravimetric moisture **Moist** (or **Dielectric-Constant Epsilon**) is shown. On the x-axis depending on the calibration curve the associated radar time **tp** in picoseconds is shown. With the software **PICO-CONFIG** the radar time **tp** is shown on the screen parallel to the moisture value **Moist**.







Calibration functions Trime-PICO64



9.11 Creating a linear Calibration Curve for a specific Material

The calibration curves Cal1 to Cal15 can be easily created or adapted for a specific soil with the help of PICO-CONFIG. Therefore, two measurement points need to be identified with the probe. Point P1 at dried soil and point P2 at moist soil where the points P1 and P2 should be far enough apart to get a best possible calibration curve. The moisture content of the soil at point P1 and P2 can be determined with laboratory measurement methods (oven drying). It is to consider that sufficient material is measured to get a representative value.

Under the menu **"Calibration**" and the window **"Material Property Calibration**" the calibration curves CAL1 to Cal15 which are stored in the PICO probe are loaded and displayed on the screen (takes max. 1 minute). With the mouse pointer individual calibration curves can be tested with the PICO-probe by activating the button **"Set Active Calib"**. The measurement of the moisture value **Moist** with the associated radar time **tp** at point P1 and P2 is started using the program PICO-CONFIG in the sub menu **"Test"** and **"Test in Mode A"**.

Step 1: The radar pulse time **tp** of the probe is measured with dried soil. It is recommended to detect multiple measurement values for finding a best average value for **tp**. The result is the first calibration point P1 (e.g. 70/0). I.e. 70ps (picoseconds) of the radar pulse time **tp** corresponds to 0% moisture content of the material. But it would be also possible to use a higher point P1' (e.g. 190/7) where a **tp** of 190ps corresponds to a moisture content of 7%. The gravimetric or volumetric moisture content of the material, e.g. 7% has to be determined with laboratory measurement methods (oven drying).

Step 2: The radar pulse time **tp** of the probe is measured with moist material. Again, it is recommended to detect multiple measurement values of **tp** for finding a best average value. The result is the second calibration point P2 with X2/Y2 (e.g. 500/25). I.e. **tp** of 500ps corresponds to 25% moisture content. The gravimetric or volumetric moisture content of the material, e.g. 25% has to be determined with laboratory measurement methods (oven drying).

Step 3: With the two calibration points P1 and P2, the calibration coefficients m0 and m1 can be determined for the specific soil (see next page).

Step 4: The coefficients m1 = 0.0581 and m0 = -4.05 (see next page) for the calibration curve Cal15 can be entered directly by hand and are stored in the probe by pressing the button "**Set**". The name of the calibration curve can also be entered by hand. The selected calibration curve (e.g. Cal15) which is activated after switching on the probes power supply will be adjusted with the button "**Set Default Calib**".

Attention: Use "dot" as separator (0.0581), not comma !

9.12 Nonlinear calibration curves

PICO probes can also work with non-linear calibration curves with polynomials up to 5th grade. Therefore it is necessary to calibrate with 4...8 different calibration points. To calculate nonlinear coefficients for polynomials up to 5th grade, the software tool TRIME-WinCal from IMKO can be used (on request). It is also possible to use any mathematical program like MATLAB for finding a best possible nonlinear calibration curve with suitable coefficient parameters m0 to m5.

The following diagram shows a sample calculation for a linear calibration curve with the coefficients m0 and m1 for a specific soil or material.





The coefficient m1 is calculated from the slope of the curve Cal15:

Coefficient $m_1 = \frac{\text{Delta y}}{\text{Delta x}} = \frac{25 - 0}{500 - 70} = 0,0581$

The coefficient m_0 is the offset on the y-axis at x = 0:

Coefficient m₀ = Y2 - (m₁ * X2) = 25 - (0,0581 * 500) = -4,05
10 Electrical Conductivity with TRIME Probes

A reliable, practical and easy-to-use Determination of Soil Conductivity and Salt Content with TRIME Probes

IMKO's TRIME TDR-probes can now report soil EC as standard simultaneously with soil moisture content percentage. A manual conversion based on researched curves for different soil types enables the user to derive a soil EC expressed in mg/I TDS (total dissolved salts).

TRIME TDR probes can do this very accurately because:

TRIME probes measure conductivity with the same large soil volume as it will be used for the TDR moisture measurement. The contact of the probe rods inside the soil is far less critical as with "gal-vanic" EC probes with a point to point measurement where even small air gaps lead to significant deviations.

TRIME probes use coated and therefore isolated rods which guarantee the non-appearance of galvanic accumulation along the rods allowing for long-run installations over many years. Unisolated rods means there is a risk of galvanic reactions and possible influence on the sensor's reading with serious problems when the probes must be removed from larger depths due to a rod cleaning.

TRIME probes measure moisture and conductivity very precisely at a frequency of 1GHz with a better and more exact separation of moisture and conductivity in comparison to capacitive probes with lower frequencies. This means that in practice, a reliable determination of the pore water conductivity EC_w and respectively TDS (mg of salt per liter water) is possible <u>at different</u> moisture levels.

All TRIME probes work with a concurrently basic calibration for moisture and conductivity. This allows a check of the limits of saline stress in soils according to standards of FAO2006 for specific soils.







10.1 The analysis of Soils for Electrical Conductivity EC_{TRIME}

Notes and Disclaimer: The analysis of electrical conductivity is an immensely complex subject. This resume is not intended as a study document in soil science nor does it discuss the technical complexities of the devices used.

For agricultural and horticultural soils, the measurement of Electrical Conductivity is an immensely important measurement. Electrical Conductivity measures the amount of total dissolved salts (TDS) or total dissolved ions in water. To complicate matters, some ions such as Sodium and Chloride will contribute more to EC than others such as Phosphorus and Potassium.

Plants require nutrients such as Nitrogen, Phosphorus, Potassium, Magnesium in large quantities hence they are called major nutrients and also smaller amounts of elements such as Iron, Manganese, Molybdenum and these are called micro nutrients or sometimes referred to as trace metals. Fertilisers are supplied to plants as compounds for example Ammonium Nitrate which supplies Nitrogen in the form of Nitrate or Ammonium. Micro-organisms will break down these compounds so they are more readily available for uptake by the plants. Levels of some ions such as Chlorides are less desirable and in great quantities can be harmful to plant growth. The quantity of ions or salts in a soil is of huge importance. Too much or too few nutrients will create a restriction in plant growth.

Measurement of EC in water is relatively straightforward. An EC probe (usually platinum) is inserted in the water and a reading in mS/cm is reported on the meter's display. This is relatively easy because water is a homogenous medium. Soil on the other hand is not and this has caused great difficulties when trying to measure its conductivity. To work around this, the analysis on agricultural and horticultural soils has been carried out by mixing a volume of soil and a volume of water and measuring the EC on the suspension or filtered extract. Different countries and different regulatory bodies have specified methodologies and in general results have been good enough to use for fertilisation recommendations and programmes. This analysis was generally carried out by a laboratory and whereas results were and are reliable the method is slow, time consuming and expensive.

Over the years manufacturers have tried to come up with solutions whereby an instrument could be used directly in soils without having to resort to laboratories. The EC reported by these instruments is referred to soil bulk electrical conductivity. Whereas there has been some success, there are so many influential variables such as temperature, soil moisture and granular composition that results have not been adequate for reliable fertilisation studies. Mostly soil bulk conductivity was of academic rather than practical interest.

IMKO has studied the subject in detail and has come up with a breakthrough. By using coated rods and measuring over the length of the probes, all TRIME probes can now accurately report what IMKO calls EC_{TRIME} . This measurement takes account of soil moisture and conductivity by volume. Because soil moisture is so important in the calculation of EC, all different TRIME probes now incorporate TDR calibration curves for a selection of soils. Special graphs have been constructed so that the user can convert the EC_{TRIME} reading to grams/litre of dissolved salt. So far curves are available for sandy and loam soils and it is intended to produce a handful of curves to cover most situations. At this moment in time, conversion of EC_{TRIME} to mg/I TDS is done manually.

IMKO hopes that this breakthrough whereby a TDR instrument can be used to derive a true soil EC measurement will become the new standard for soil fertilisation analyses.



10.2 EC_{TRIME} Measurement Range for TRIME Probes

All PICO probes work with a concurrently basic calibration for moisture and conductivity. This allows a check of the limits of saline soils according to standards of FAO2006 for specific soils.

Measurement Range Conductivity for PICO64, PICO32, T3-IPH, PICO-PROFILE

 $0...10dS/m EC_{TRIME}$ (here is, depending on moisture content $EC_w > 20dS/m$)

All TRIME probes have a linear EC_{TRIME} calibration curve with two calibration points. This ensures an equal behaviour of a TRIME probe type (e.g. PICO32).

Calibration Point1: Dry glass beads = OdS/m EC_{TRIME}

Calibration Point2: Saturated glass beads with water $EC_w 5dS/m = 5dS/m EC_{TRIME}$

With dry glass beads (not air!) it is ensured that the conductivity range of dry soils starts with nearly 0dS/m. The calibration at point2 was performed with glass beads and water which was adjusted with salt to a pore water conductivity of $EC_w = 5$ dS/m at a temperature of 23°C.

10.3 Measurement of Pore-Water-Conductivity or Salt-Content of Soils

The possible salt stress depends on soil type and plant species and can be roughly classified:

Slight salinization:	1-3g of salt per liter
Average salinity:	3-5g salt / liter
Strong salinity:	5-10g salt / liter
Very high salinity:	> 10 g salt / liter

The conversion of the salt content in grams of salt per liter water (TDS Total-Dissolved-Salts) into the pore water conductivity of EC_w happens with:

Salt content in grams of salt per liter of water TDS = $0.64 * EC_w$ An EC_w = 5mS/cm result thus with 3.2 grams of salt per liter (pore) water EC_w can be determined with a laboratory conductivity meter. 1dS/m = 1mS/cm

A determination of the pore water conductivity or salinity of soils and sand at different moisture values is possible by using a Moisture/ EC_{TRIME} diagram for the PICO32 or PICO64. With measuring moisture and EC_{TRIME} , the pore water conductivity (or TDS salt content in grams salt per liter) of the soil can be determined by using the individual pore water curves.

The below displayed Moisture/ EC_{TRIME} charts were created for sand and a loamy soil.

We recommend the creation of a separate Moisture/ EC_{TRIME} diagram for the intended soil in the designated ranges of salinity.



The following diagram illustrates curve progressions of pore-water with four different salinity contents at different moisture contents for the PICO32.

Sand1 was mixed with 0,5dS/m pore water, sand2 mixed with water which had a conductivity of 5dS/m, sand3 with 10dS/m and sand4 with 20dS/m water. The PICO32 determines different EC_{TRIME} conductivities in the different sands at one moisture level. The moisture range with the best resolution for the EC_{TRIME} measurement lies between 10% and 30%, depending on the salt content.



How to determine the salt content in sand

With both parameters, the moisture value and EC_{TRIME} it is possible to determine the salt content in sand. With a moisture content of e.g. 17% and an EC_{TRIME} of 4,6dS/m, the intersection point lies on the blue curve which represents a salt content of 6,4g salt per liter water in the sand. With 30% moisture content and an EC_{TRIME} of 5,5dS/m the salt content would be 3,2g salt per liter water in the sand.

This sand diagram shows four pore water curves. If intersection points lies between this four curves (e.g. the yellow curve), the result of the salt content has to be interpolated.

The following diagram illustrates curve progressions of EC_{TRIME} for PICO32 for a loamy soil at different salt contents. The PICO32 determines different EC_{TRIME} conductivities in different soils at one moisture level. The moisture range with the best resolution for the EC_{TRIME} measurement for a loamy soil is between 25% and 35%.



How to determine the salt content in loamy soil

With both parameters, the moisture value and EC_{TRIME} it is possible to determine the salt content in loamy soil. With a moisture content of e.g. 28% and an EC_{TRIME} of 3dS/m, the intersection point lies on the red curve which represents a salt content of 0.38g salt per liter water in the soil. With 28% moisture content and an EC_{TRIME} of 5,5dS/m the salt content would be 7g salt per liter water in the soil.

This diagram shows five pore water curves. If intersection points lies between this five curves, the result of the salt content has to be interpolated.



10.4 Creation of a Moisture/ EC_{TRIME} Diagram for a particular Soil

The following description shows a possible way to create a Moisture/ EC_{TRIME} diagram for a particular soil:

Take a soil sample and determine the salt content with a laboratory method. (e.g. determine EC_w of the soil/water saturation extract).

Dry the soil sample to 0% moisture. Cut it into small pieces and grind the soil so that it can be moistened by hand as evenly as possible (eg with a water spray pump).

Moisten the soil evenly with a pore water solution for example 2.56g of salt per liter (equivalent to EC_w = 4dS/m) to the first measurement point (e.g 2.5% moisture). Take into account, that the soil already contains salt. I.e. were already 1g of salt per liter in the soil, you put a curve for a pore water conductivity record of 2.56 g + 1g salt = 3.56g per liter (EC_w = 5.5 dS/m).

Compact the soil in a plastic bowl (no metal and not too small!) so that it largely corresponds to the natural density. Concerning the insertion of the PICO-probe rods for measurement, it is recommended to perform several measurements per point in order to form an average value for each moisture and EC_{TRIME} value. It is important to ensure that the measuring field is not disturbed around the rods, eg by previous punctures.

Moisten the soil more, as homogeneous as possible with the pore water solution EC_w = 4dS/m to eg 5% moisture. Again compact the soil in a plastic bowl so that it largely corresponds to the natural density. Concerning the insertion of the PICO-probe rods for measurement, it is recommended to perform several measurements per point in order to form an average value for each moisture and EC_{TRIME} value. It is important to ensure that the measuring field is not disturbed around the rods, eg by previous punctures.

Take more moisture and EC_{TRIME} values, up to soil saturation.

Create the pore-water curve for example for 3.56g salt/liter (EC_w =5.5dS/m) with the added moisture and EC-TRIME values.

Take up several curves for different pore water contents.

If the soil sample is already too heavily loaded with salt, a washing out of the ions would be possible, but without the lost of the fine fraction. After a subsequent drying and crushing of the soil, the Moisture/ EC_{TRIME} diagram can be established with a lower salt content. It is also possible to use distilled water for obtaining the lowest curve.

<u>Note:</u> The effort for a Moisture/ EC_{TRIME} diagram is not negligible. IMKO has the hope to establish with the EC_{TRIME} process an easier, faster, better and generally accepted determination of the salt contamination of soils. In future IMKO plans to provide or publish different Moisture/ EC_{TRIME} curves for standard soils. We expect to cover nearly most soils with about 4-5 diagrams.



11 Software Error codes

11.1 Software Errors which will be coded with 4 digits.

Code	Explanation	Measurement
0101-0108, 0301	Serial Port errors	Check port's setting or if the port has been opened. Then close and restart the program.
0201	No answer	Check the power of PICO, the serial port of PC and the connection between PC and PICO.
0202-0212	Protocol errors	Check if PICO-CONFIG's version passes PICO's version.
0302-0307	Protocol error, Parameter setting false	Check if parameters are correct and if PICO-CONFIG's version passes PICO's version.
0401	Can not find config file PICO- CONFIG.con	Look for the file in the exe path. If not found, copy the file to the path.
0501-0508	Errors in Event & MeasureMode	Restart PICO and PICO-CONFIG
0601,0602,0604, 0605,0606,0607, 0609	Operation errors in Basic Balancing	Operate correctly and try it again.
0603,0608,0610, 0611	Communication or protocol errors in Basic Balancing	Restart PICO and PICO- CONFIG.
0701,0702	Read file errors in Material Property Calibration	Check if the files are in the required path. If not, copy the files to the path or redefine the path under the menu Bus/Configuration/Material Property Calibration.
0703	Operation errors in Material Property Calibration	Operate correctly and try it again.
0704, 0705,0706	Communication or protocol errors in Material Property Calibration	Restart PICO and PICO- CONFIG.
0801-0805,0901	Operation errors in Calibration IDs and Names	Operate correctly and try it again.
1001,1101,1102	Read file errors in Calibration IDs and Names	Check if the files are in the required path. If not, copy the files to the path or redefine the path under the menu Bus/Configuration/Material Property Calibration.
1201, 1203	Operation errors in Test	Operate correctly and try it
· · · · · · · · · · · · · · · · · · ·		



		again.
1202	Communication or protocol errors in Test	Restart PICO and PICO- CONFIG.
1301	Write file error in Test	Check file path and try it again. If failed, restart PICO-CONFIG.
4001-4002, 4101-4106,4201	Read file or write file errors	Check the files and try again.
4301-4303	Intenal calculating errors	Restart PICO-CONFIG and try again.Otherwise contact IMKO.



11.2 PICO Errors (Firmware errors)

The errors come from the firmware, from 1 to 255.

Code	Explanation	Measurement
1-19	The serial communication errors due to incorrect telegram, baud rate, timing etc.	Power off, power on PICO and try it again. Otherwise contact IMKO.
20-39	incorrect command number, command right or command parameters.	Power off, power on PICO and try it again. Otherwise contact IMKO.
40-49	EEPROM is defect	Power off, power on PICO and try it again. Otherwise contact IMKO.
50-59	ASIC is defect	Power off, power on PICO and try it again. Otherwise contact IMKO.
60	Power voltage is too low	Check power voltage of PICO, it is minimal 6V.
100	TDR measurement parameter is incorrectly set or the material conductivity is too high.	Adjust the TDR measurement parameters or contact IMKO
101	TDR measurement parameter is incorrectly set	Adjust the TDR measurement parameters or contact IMKO
102	ASIC is defect	Power off, power on PICO and try it again. Otherwise contact IMKO.
103	EC parameter is incorrectly set.	Power off, power on PICO and try it again. Otherwise contact IMKO.
105	Tp is out of range for the standard calibra- tion polynomial.	Check if PICO is inserted in the measured ma- terial correctly.
108	TDR measurement parameter is incorrectly set or the material conductivity is too high.	Adjust the TDR measurement parameters or contact IMKO
120-129	Internal chip problem	Power off, power on PICO and try it again. Otherwise contact IMKO.
130-199	Internal errors	Contact IMKO
200-254	Reserved	
255	The data transmission is not finished.	



12 Savety Notes

In this documentation, text points are highlighted, which require special attention.



DANGER:

The Warning Triangle with the exclamation mark warns you against personal injury or property damage.

Intended Use

Sensors and measuring systems of IMKO GmbH may only be used for the purpose described, taking into account the technical data. Misuse **and use of the e**quipment other than for its intended purpose **are not eligible.** The function and operational safety of a sensor or measuring system can only be guaranteed if the general safety precautions, national regulations and the special safety instructions in this operating manual are observed during use. The moisture sensors and measuring systems of IMKO GmbH are used to measure moisture according to the measuring purpose and measuring range defined and defined in the technical data. Only adherence to the instructions described in the manual is regarded as intended use. The manual describes the connection, use and maintenance of IMKO sensors and IMKO measuring systems. Read the manual before connecting and operating a sensor or measuring system. The manual is part of the product and must be kept close to the sensor or measuring system.



Impairment of safety

The sensor or the measuring system has been designed and tested in accordance with EN 61010 safety regulations for electronic measuring instruments and has left the factory in a safe and safe condition. If the sensor or the measuring system can no longer be operated safely, it must be put out of operation

and secured by means of marking before further commissioning. In case of doubt, the sensor or the measuring system must be sent to the manufacturer or his contractual partner for repair or maintenance.



Modifications

For safety reasons, it is not permitted to carry out any modifications or modifications to the sensor or the measuring system without the consent of the manufacturer. The opening of the sensor or hand-held meter, adjustment and repair work, as well as all maintenance work other than the work described in the

manual may only be carried out by a specialist authorized by IMKO. The sensor or the measuring system must be disconnected from the power supply before installation or maintenance work. Do not open or repair the hand-held unit and the power supply!



Hazard Warnings

Danger due to improper operation. The sensor or the measuring system may only be operated by instructed personnel. The operating personnel must have read and understood the operating instructions.





Danger by electricity

The hand-held meter must not be immersed in water or other liquids. The sensor is insensitive to moisture contained in the typically measured products. Only connect the hand-held meter to a properly installed outlet with the supplied voltage supply cable, the voltage of which corresponds to the tech-

nical data. Make sure that the power outlet is well accessible, so that you can unplug the power supply quickly if necessary. Use only the adapter that is suitable for your outlet.

Only operate the meter with the supplied original accessories. If you need additional accessories or replacement, please contact the manufacturer.

Do not use the meter in following case:

- if the measuring instrument, sensor, plug-in power supply or accessories are damaged,
- the sensor or the measuring system does not operate as intended,
- the power cord or plug is damaged,
- the sensor or the measuring system has fallen down.

Unplug the power supply from the wall outlet in following case:

- if you do not use the sensor or the measuring system for an extended period of time,
- before cleaning, unpacking or changing the sensor or the measuring system,
- if you are working inside the sensor or measuring instrument, e.g. connect devices,
- if a fault occurs during operation,
- during thunderstorms.



Caution - Property damage

Ensure that there is a sufficient distance to strong heat sources such as heating plates, heating pipes. Disconnect the sensor or handheld device from other devices before relocating or transporting it. Disconnect the connectors on the device.

Do not use aggressive chemical cleaning agents, scouring agents, hard sponges or the like.



Precise Moisture Measurement

in hydrology, forestry, agriculture, environmental and earth science, civil engineering, as well as individual applications!